

EC CERTIFICATE

Number: 2188436CE01

Full Quality Assurance System

Directive 90/385/EEC on Active Implantable Medical Devices, Annex 2 excluding (4)
(Other devices than custom made or intended for clinical investigation)

Manufacturer:

Implandata Ophthalmic Products GmbH

Kokenstrasse 5
30159 Hannover
Germany

For the product category(ies)

Intraocular pressure sensor for measurement of intraocular pressure in patients with primary open angle glaucoma

DEKRA grants the right to use the EC Notified Body Identification Number illustrated below to accompany the CE Marking of Conformity on the products concerned conforming to the required Technical Documentation and meeting the provisions of the EC-Directive which apply to them:

0344

Documents, that form the basis of this certificate:

Certification Notice 2188436CN, initially dated 8 November 2016 **Addendum, initially dated 24 May 2017**

DEKRA hereby declares that the above mentioned manufacturer fulfils the relevant provisions of 'Besluit Actieve Implantaten', the Dutch transposition of the Directive 90/385/EEC of 20 July 1990 concerning Active implantable medical devices, including all subsequent amendments. The manufacturer has implemented a quality assurance system for design, manufacture and final inspection for the above mentioned product category in accordance to the provisions of Annex 2 of Council Directive 90/385/EEC of 20 July 1990 and is subject to periodical surveillance. For placing on the market of Active implantable medical devices an additional EC design examination certificate according to Annex 2 (4) is mandatory.

The necessary information related to the quality management system of the manufacturer, including facilities and the reference to the relevant documentation, of the products concerned and the assessments performed, are stated in the Certification Notice which forms an integrative part of this certificate.

This certificate is valid until: 1 June 2020
Issued for the first time: 24 May 2017

DEKRA Certification B.V.



drs. G.J. Zoetbrood
Managing Director



ing. A.A.M. Laan
Certification Manager

© Integral publication of this certificate and adjoining reports is allowed

DEKRA Certification B.V. is Notified Body with ID no 0344

DEKRA Certification B.V. Meander 1051, 6825 MJ Arnhem P.O. Box 5185, 6802 ED Arnhem, The Netherlands
T +31 88 96 83000 F +31 88 96 83100 www.dekra-certification.com Company registration 09085396

ADDENDUM

Belonging to certificate: 2188436CE01

1/1

CE MARKING OF CONFORMITY ACTIVE IMPLANTABLE MEDICAL DEVICES

Intraocular pressure sensor for measurement of intraocular pressure in patients with primary open angle glaucoma

Issued to:

Impladata Ophthalmic Products GmbH
Kokenstrasse 5
30159 Hannover
Germany

This certificate covers the following product(s):

EYEMATE-IO, intraocular pressure sensor for measurement of intraocular pressure in patients with primary open angle glaucoma

Initial date: 24 May 2017

DEKRA Certification B.V.

A blue ink signature of drs. G.J. Zoetbrood, consisting of stylized cursive letters.

drs. G.J. Zoetbrood
Managing Director

A blue ink signature of ing. A.A.M. Laan, consisting of stylized cursive letters.

ing. A.A.M. Laan
Certification Manager

© Integral publication of this certificate and adjoining reports is allowed

DEKRA Certification B.V. is Notified Body with ID no 0344

DEKRA Certification B.V. Meander 1051, 6825 MJ Arnhem P.O. Box 5185, 6802 ED Arnhem, The Netherlands
T +31 88 96 83000 F +31 88 96 83100 www.dekra-certification.com Company registration 09085396